

Hemp Proficiency Testing

All Labs Trueness

OL22SEP-1 (oil)



Sample: OL22SEP-1 (oil)

All Labs Trueness

Issue Date: November 18, 2022

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
			Δ9-THC (%AR)									
Δ9-THC (%AR)	001.60	NIR	187	0.02	0.03	0.03	0.0267	-2.95	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	184	0.0554	0.0548	0.056	0.0554	-1.68	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.99	Other	183	0.0578	0.0636	0.0717	0.0644	-1.28	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	204	0.0673	0.07	0.0698	0.069	-1.07	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.03	AOAC 2018.11, M	139	0.0818	0.0621	0.07	0.0713	-0.97	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	191	0.0807	0.0689	0.0805	0.0767	-0.73	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	197	0.0849	0.081	0.0853	0.0837	-0.42	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	192	0.0944	0.088	0.0748	0.0857	-0.33	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	216	0.0855	0.0871	0.0868	0.0865	-0.3	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.30	Other LC-MS	102	0.1032	0.0694	0.0957	0.0894	-0.17	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.01	AOAC 2018.10	183	0.0886	0.09	0.0895	0.0894	-0.17	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	200	0.089	0.089	0.089	0.089	-0.19	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.02	AOAC 2018.11, U	125	0.0913	0.0883	0.0911	0.0902	-0.13	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.30	Other LC-MS	149	0.097	0.085	0.092	0.0913	-0.08	0.0932	26	0.0225	0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
Δ9-THC (%AR)	001.10	Other LC-UV	194	0.0947	0.0952	0.0828	0.0909	-0.1	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	106	0.09	0.1	0.09	0.0933	0	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	149	0.096	0.096	0.095	0.0957	0.11	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	196	0.1004	0.1024	0.1009	0.101	0.35	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.30	Other LC-MS	190	0.101			0.101					1
Δ9-THC (%AR)	001.10	Other LC-UV	189	0.1041	0.1081	0.109	0.107	0.61	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	212	0.1057	0.1145	0.1029	0.108	0.64	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	155	0.111	0.121	0.105	0.112	0.85	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.02	AOAC 2018.11, U	120	0.1118	0.1152	0.1193	0.115	0.98	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.10	Other LC-UV	162	0.1131	0.1143	0.1187	0.115	0.98	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.02	AOAC 2018.11, U	199	0.1265	0.1175	0.133	0.126	1.44	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.01	AOAC 2018.10	170	0.1243	0.1135	0.1437	0.127	1.51	0.0932	26	0.0225	0
Δ9-THC (%AR)	001.30	Other LC-MS	122	0.1274	0.1275	0.1269	0.127	1.51	0.0932	26	0.0225	0

Δ9-THCA (%AR)

Δ9-THCA (%AR)	002.30	Other LC-MS	190	<0.0010			<0.0010					1
Δ9-THCA (%AR)	002.02	AOAC 2018.11, U	120	<0.0025	<0.0025	<0.0025	<0.0025					1
Δ9-THCA (%AR)	002.30	Other LC-MS	149	<0.0040	<0.0040	<0.0040	<0.0040					1
Δ9-THCA (%AR)	002.10	Other LC-UV	197	<0.0094	<0.0098	<0.0092	<0.0094					1
Δ9-THCA (%AR)	002.10	Other LC-UV	216	<0.0099	<0.0091	<0.0103	<0.0099					1
Δ9-THCA (%AR)	002.10	Other LC-UV	162	<0.0100	<0.0100	<0.0100	<0.0100					1
Δ9-THCA (%AR)	002.10	Other LC-UV	106	<0.0100	<0.0100	<0.0100	<0.0100					1

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
Δ9-THCA (%AR)	002.10	Other LC-UV	191	<0.0100	<0.0100	<0.0100	<0.0100					1
Δ9-THCA (%AR)	002.03	AOAC 2018.11, M	139	<0.0125	<0.0125	<0.0125	<0.0125					1
Δ9-THCA (%AR)	002.30	Other LC-MS	102	<0.0194	<0.0188	<0.0187	<0.0194					1
Δ9-THCA (%AR)	002.10	Other LC-UV	196	<0.0210	<0.0210	<0.0210	<0.0210					1
Δ9-THCA (%AR)	002.10	Other LC-UV	194	<0.0270	<0.0270	<0.0270	<0.0270					1
Δ9-THCA (%AR)	002.10	Other LC-UV	149	<0.0370	<0.0370	<0.0400	<0.0370					1
Δ9-THCA (%AR)	002.02	AOAC 2018.11, U	199	<0.0500	<0.0500	<0.0500	<0.0500					1
Δ9-THCA (%AR)	002.30	Other LC-MS	122	<0.0500	<0.0500	<0.0500	<0.0500					1
Δ9-THCA (%AR)	002.10	Other LC-UV	184	<0.1000	<0.1000	<0.1000	<0.1000					1
Δ9-THCA (%AR)	002.01	AOAC 2018.10	170	<0.1300	<0.1300	<0.1300	<0.1300					1
Δ9-THCA (%AR)	002.01	AOAC 2018.10	183	0	0	0	0					2
Δ9-THCA (%AR)	002.02	AOAC 2018.11, U	125	0.002	0.0021	0.0025	0.0022	-1.05	0.0309	6	0.0274	0
Δ9-THCA (%AR)	002.60	NIR	187	0.01	0.02	0.01	0.0133	-0.64	0.0309	6	0.0274	0
Δ9-THCA (%AR)	002.10	Other LC-UV	200	0.022	0.022	0.022	0.022	-0.32	0.0309	6	0.0274	0
Δ9-THCA (%AR)	002.10	Other LC-UV	189	0.0341	0.0377	0.0387	0.0368	0.22	0.0309	6	0.0274	0
Δ9-THCA (%AR)	002.10	Other LC-UV	155	0.05	0.03	0	0.04	0.33	0.0309	6	0.0274	0
Δ9-THCA (%AR)	002.10	Other LC-UV	192	0.0729	0.0647	0.0748	0.0708	1.46	0.0309	6	0.0274	0

CBD (%AR)

CBD (%AR)	003.10	Other LC-UV	191	4.6798	3.1997	3.3241	3.73	-3.67	5.7535	24	0.5499	0
CBD (%AR)	003.03	AOAC 2018.11, M	139	5.1067	4.0449	4.3818	4.51	-2.26	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	200	5.128	4.946	4.977	5.02	-1.34	5.7535	24	0.5499	0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
CBD (%AR)	003.10	Other LC-UV	189	5.0957	5.1238	5.0945	5.1	-1.18	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	192	5.4912	5.2414	4.7246	5.15	-1.09	5.7535	24	0.5499	0
CBD (%AR)	003.99	Other	212	5.4674	5.4764	5.4158	5.45	-0.55	5.7535	24	0.5499	0
CBD (%AR)	003.01	AOAC 2018.10	183	5.5916	5.5525	5.5681	5.57	-0.33	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	194	5.5474	5.5454	5.6459	5.58	-0.32	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	184	5.736	5.481	5.631	5.62	-0.25	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	149	5.655	5.521	5.816	5.66	-0.16	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	155	5.67	5.74	5.71	5.71	-0.09	5.7535	24	0.5499	0
CBD (%AR)	003.01	AOAC 2018.10	170	5.6788	5.7715	5.6892	5.71	-0.07	5.7535	24	0.5499	0
CBD (%AR)	003.02	AOAC 2018.11, U	125	5.83	5.63	5.69	5.72	-0.07	5.7535	24	0.5499	0
CBD (%AR)	003.30	Other LC-MS	190	5.916			5.92					1
CBD (%AR)	003.10	Other LC-UV	204	5.9638	5.9321	5.9633	5.95	0.36	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	196	6.0616	6.0048	5.7996	5.96	0.37	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	106	6.04	5.96	6.01	6	0.45	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	216	5.8788	6.1045	6.0279	6	0.45	5.7535	24	0.5499	0
CBD (%AR)	003.30	Other LC-MS	102	5.9804	6.2777	5.7592	6.01	0.46	5.7535	24	0.5499	0
CBD (%AR)	003.02	AOAC 2018.11, U	199	5.954	5.959	6.135	6.02	0.48	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	162	6.1554	6.3674	6.1204	6.21	0.84	5.7535	24	0.5499	0
CBD (%AR)	003.10	Other LC-UV	197	6.23	6.2888	6.3578	6.29	0.98	5.7535	24	0.5499	0
CBD (%AR)	003.30	Other LC-MS	149	6.427	6.31	6.262	6.33	1.05	5.7535	24	0.5499	0
CBD (%AR)	003.60	NIR	187	6.62	6.73	6.5	6.62	1.57	5.7535	24	0.5499	0
CBD (%AR)	003.30	Other LC-MS	122	6.8498	6.8495	6.8501	6.85	1.99	5.7535	24	0.5499	0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
			CBDA (%AR)									
CBDA (%AR)	004.30	Other LC-MS	149	<0.0040	<0.0040	<0.0040	<0.0040					1
CBDA (%AR)	004.10	Other LC-UV	197	<0.0094	<0.0098	<0.0092	<0.0094					1
CBDA (%AR)	004.10	Other LC-UV	216	<0.0099	<0.0091	<0.0103	<0.0099					1
CBDA (%AR)	004.10	Other LC-UV	106	<0.0100	<0.0100	<0.0100	<0.0100					1
CBDA (%AR)	004.10	Other LC-UV	162	<0.0100	<0.0100	<0.0100	<0.0100					1
CBDA (%AR)	004.03	AOAC 2018.11, M	139	<0.0125	<0.0125	<0.0125	<0.0125					1
CBDA (%AR)	004.10	Other LC-UV	194	<0.0270	<0.0270	<0.0270	<0.0270					1
CBDA (%AR)	004.10	Other LC-UV	149	<0.0370	<0.0370	<0.0400	<0.0370					1
CBDA (%AR)	004.02	AOAC 2018.11, U	199	<0.0500	<0.0500	<0.0500	<0.0500					1
CBDA (%AR)	004.30	Other LC-MS	122	<0.0500	<0.0500	<0.0500	<0.0500					1
CBDA (%AR)	004.01	AOAC 2018.10	170	<0.1300	<0.1300	<0.1300	<0.1300					1
CBDA (%AR)	004.01	AOAC 2018.10	183	0	0	0	0					2
CBDA (%AR)	004.10	Other LC-UV	200	0.004	0.004	0.004	0.004	-1.05	0.0316	11	0.0263	0
CBDA (%AR)	004.02	AOAC 2018.11, U	125	0.005	0.0055	0.0062	0.0056	-0.99	0.0316	11	0.0263	0
CBDA (%AR)	004.10	Other LC-UV	189	0.0044	0.0098	0.0106	0.0083	-0.89	0.0316	11	0.0263	0
CBDA (%AR)	004.10	Other LC-UV	192	0.0236	0.0252	0.0218	0.0235	-0.31	0.0316	11	0.0263	0
CBDA (%AR)	004.10	Other LC-UV	191	0.0251	0.0232	0.0267	0.025	-0.25	0.0316	11	0.0263	0
CBDA (%AR)	004.10	Other LC-UV	212	0.0266	0.0294	0	0.028	-0.14	0.0316	11	0.0263	0
CBDA (%AR)	004.10	Other LC-UV	196	0.0325	0.032	0.0322	0.0322	0.02	0.0316	11	0.0263	0
CBDA (%AR)	004.10	Other LC-UV	204	0.0351	0.0328	0.0405	0.0361	0.17	0.0316	11	0.0263	0
CBDA (%AR)	004.10	Other LC-UV	184	0.0438	0.0426	0.0431	0.0432	0.44	0.0316	11	0.0263	0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
CBDA (%AR)	004.10	Other LC-UV	155	0.079	0.075	0.077	0.077	1.73	0.0316	11	0.0263	0
CBDA (%AR)	004.60	NIR	187	0.15	0.19	0.23	0.19	6.02	0.0316	11	0.0263	0

CBN (%AR)

CBN (%AR)	005.20	Other LC-MS	102	<0.0484	<0.0470	<0.0467	<0.0484					1
CBN (%AR)	005.02	AOAC 2018.11, U	199	<0.0500	<0.0500	<0.0500	<0.0500					1
CBN (%AR)	005.03	AOAC 2018.11, M	139	0.0338	0.0304	0.0275	0.0306	-1.58	0.0424	17	0.0074	0
CBN (%AR)	005.02	AOAC 2018.11, U	107	0.032	0.029	0.045	0.0353	-0.94	0.0424	17	0.0074	0
CBN (%AR)	005.10	Other LC-UV	194	0.0368	0.0368	0.0309	0.0348	-1.01	0.0424	17	0.0074	0
CBN (%AR)	005.10	Other LC-UV	197	0.0362	0.0363	0.0371	0.0365	-0.78	0.0424	17	0.0074	0
CBN (%AR)	005.10	Other LC-UV	216	0.0372	0.0382	0.0384	0.0379	-0.6	0.0424	17	0.0074	0
CBN (%AR)	005.10	Other LC-UV	149	0.039	<0.0370	<0.0400	0.0387					1
CBN (%AR)	005.01	AOAC 2018.10	183	0.0397	0.039	0.0396	0.0394	-0.39	0.0424	17	0.0074	0
CBN (%AR)	005.10	Other LC-UV	200	0.04	0.04	0.04	0.04	-0.32	0.0424	17	0.0074	0
CBN (%AR)	005.10	Other LC-UV	189	0.0404	0.0411	0.0412	0.0409	-0.2	0.0424	17	0.0074	0
CBN (%AR)	005.20	Other LC-MS	149	0.044	0.04	0.043	0.0423	0	0.0424	17	0.0074	0
CBN (%AR)	005.10	Other LC-UV	204	0.0402	0.0435	0.0427	0.0421	-0.03	0.0424	17	0.0074	0
CBN (%AR)	005.10	Other LC-UV	196	0.0427	0.043	0.0423	0.0427	0.04	0.0424	17	0.0074	0
CBN (%AR)	005.02	AOAC 2018.11, U	125	0.0444	0.0418	0.047	0.0444	0.27	0.0424	17	0.0074	0
CBN (%AR)	005.10	Other LC-UV	162	0.0442	0.0462	0.0448	0.0451	0.36	0.0424	17	0.0074	0
CBN (%AR)	005.20	Other LC-MS	190	0.048			0.048					1
CBN (%AR)	005.10	Other LC-UV	212	0.0501	0.0472	0.0484	0.0486	0.83	0.0424	17	0.0074	0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
CBN (%AR)	005.10	Other LC-UV	192	0.0589	0.0538	0.0435	0.0521	1.3	0.0424	17	0.0074	0
CBN (%AR)	005.10	Other LC-UV	106	0.05	0.06	0.05	0.0533	1.47	0.0424	17	0.0074	0
CBN (%AR)	005.30	GC-FID	138	0.0554	0.0526	0.0618	0.0566	1.91	0.0424	17	0.0074	0
Total Δ9-THC (%AR)												
Total Δ9-THC (%AR)	006.60	NIR	187	0.029	0.048	0.039	0.0387	-2.38	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.99	Other	183	0.0578	0.0636	0.0717	0.0644	-1.37	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	204	0.0673	0.07	0.0698	0.069	-1.19	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.03	AOAC 2018.11, M	139	0.0818	0.0621	0.07	0.0713	-1.1	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	191	0.0807	0.0689	0.0805	0.0767	-0.89	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	197	0.0849	0.081	0.0853	0.0837	-0.61	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	216	0.0855	0.0871	0.0868	0.0865	-0.5	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.02	AOAC 2018.11, U	107	0.078	0.083	0.102	0.0877	-0.46	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.01	AOAC 2018.10	183	0.0886	0.09	0.0895	0.0894	-0.39	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.30	Other LC-MS	102	0.1032	0.0694	0.0957	0.0894	-0.39	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.30	Other LC-MS	149	0.097	0.085	0.092	0.0913	-0.31	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	194	0.0947	0.0952	0.0826	0.0908	-0.33	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.02	AOAC 2018.11, U	125	0.0931	0.0902	0.0933	0.0922	-0.28	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	149	0.096	0.096	0.095	0.0957	-0.14	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.30	Other LC-MS	190	0.101			0.101					1
Total Δ9-THC (%AR)	006.10	Other LC-UV	196	0.1004	0.1024	0.1009	0.101	0.08	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.40	GC-FID	138	0.1018	0.0992	0.1045	0.102	0.1	0.0993	26	0.0255	0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
Total Δ9-THC (%AR)	006.10	Other LC-UV	212	0.1057	0.1145	0.1029	0.108	0.33	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	200	0.108	0.111	0.108	0.109	0.38	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.02	AOAC 2018.11, U	120	0.1118	0.1152	0.1193	0.115	0.63	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	162	0.1131	0.1143	0.1187	0.115	0.63	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.40	GC-FID	169	0.1193	0.111	0.1175	0.116	0.65	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.30	Other LC-MS	122	0.1274	0.1275	0.1269	0.127	1.1	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.02	AOAC 2018.11, U	199	0.1309	0.1175	0.133	0.127	1.09	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	155	0.156	0.148	0.105	0.136	1.45	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	189	0.134	0.1411	0.1429	0.139	1.57	0.0993	26	0.0255	0
Total Δ9-THC (%AR)	006.10	Other LC-UV	192	0.1583	0.1447	0.1396	0.148	1.89	0.0993	26	0.0255	0

Total CBD (%AR)

Total CBD (%AR)	007.40	GC-FID	169	3.6661	3.337	3.3931	3.47	-2.54	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	191	4.7018	3.22	3.3475	3.76	-2.21	5.6662	23	0.8652	0
Total CBD (%AR)	007.03	AOAC 2018.11, M	139	5.1067	4.0449	4.3818	4.51	-1.34	5.6662	23	0.8652	0
Total CBD (%AR)	007.02	AOAC 2018.11, U	107	4.491	4.237	5.03	4.59	-1.25	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	200	5.132	4.95	4.981	5.02	-0.75	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	189	5.0995	5.1323	5.1038	5.11	-0.64	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	192	5.5119	5.2635	4.7437	5.17	-0.57	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	212	5.4907	5.5022	5.4158	5.47	-0.23	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	194	5.5474	5.5454	5.5459	5.55	-0.14	5.6662	23	0.8652	0
Total CBD (%AR)	007.01	AOAC 2018.10	183	5.5916	5.5525	5.5681	5.57	-0.11	5.6662	23	0.8652	0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
Total CBD (%AR)	007.10	Other LC-UV	149	5.655	5.521	5.816	5.66	0	5.6662	23	0.8652	0
Total CBD (%AR)	007.02	AOAC 2018.11, U	125	5.84	5.63	5.69	5.72	0.06	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	155	5.74	5.8	5.78	5.77	0.12	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	196	6.09	6.0329	5.8278	5.98	0.37	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	204	5.9946	5.9921	5.9676	5.98	0.37	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	216	5.8788	6.1045	6.0279	6	0.39	5.6662	23	0.8652	0
Total CBD (%AR)	007.02	AOAC 2018.11, U	199	5.9851	5.9888	6.1666	6.05	0.44	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	162	6.1554	6.3674	6.1204	6.21	0.63	5.6662	23	0.8652	0
Total CBD (%AR)	007.10	Other LC-UV	197	6.23	6.2888	6.3578	6.29	0.72	5.6662	23	0.8652	0
Total CBD (%AR)	007.30	Other LC-MS	149	6.427	6.31	6.262	6.33	0.77	5.6662	23	0.8652	0
Total CBD (%AR)	007.60	NIR	187	6.7	6.9	6.7	6.77	1.27	5.6662	23	0.8652	0
Total CBD (%AR)	007.30	Other LC-MS	122	6.8498	6.8495	6.8501	6.85	1.37	5.6662	23	0.8652	0
Total CBD (%AR)	007.40	GC-FID	138	6.8387	7.382	7.5153	7.25	1.83	5.6662	23	0.8652	0

CBDV (%AR)

CBDV (%AR)	008.99	Other	162	<0.0100	<0.0100	<0.0100	<0.0100					1
CBDV (%AR)	008.99	Other	197	<0.0142	<0.0146	<0.0138	<0.0142					1
CBDV (%AR)	008.99	Other	216	<0.0148	<0.0136	<0.0154	<0.0148					1
CBDV (%AR)	008.99	Other	196	<0.0210	<0.0210	<0.0210	<0.0210					1
CBDV (%AR)	008.99	Other	189	0.018	0.017	0.0182	0.0177	-0.89	0.0248	7	0.0079	0
CBDV (%AR)	008.02	AOAC 2018.11, U	183	0.0208	0.0213	0.0206	0.0209	-0.49	0.0248	7	0.0079	0
CBDV (%AR)	008.99	Other	149	0.021	0.02	0.021	0.0207	-0.52	0.0248	7	0.0079	0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
CBDV (%AR)	008.99	Other	200	0.023	0.022	0.022	0.0223	-0.31	0.0248	7	0.0079	0
CBDV (%AR)	008.99	Other	194	0.0238	0.0246	0.0191	0.0225	-0.29	0.0248	7	0.0079	0
CBDV (%AR)	008.99	Other	106	0.02	0.04	0.04	0.0333	1.09	0.0248	7	0.0079	0
CBDV (%AR)	008.99	Other	212	0.0363	0.0358	0.0355	0.0359	1.41	0.0248	7	0.0079	0

CBDVA (%AR)

CBDVA (%AR)	009.99	Other	197	<0.0094	<0.0098	<0.0092	<0.0094					1
CBDVA (%AR)	009.99	Other	216	<0.0099	<0.0091	<0.0103	<0.0099					1
CBDVA (%AR)	009.99	Other	196	<0.0210	<0.0210	<0.0210	<0.0210					1
CBDVA (%AR)	009.99	Other	149	<0.0370	<0.0370	<0.0400	<0.0370					1
CBDVA (%AR)	009.01	AOAC 2018.10	183	0	0	0	0					2

CBG (%AR)

CBG (%AR)	010.99	Other	149	0.067	0.07	0.067	0.068	-1.05	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	197	0.0715	0.0714	0.071	0.0713	-0.96	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	216	0.0721	0.0741	0.0725	0.0729	-0.91	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	194	0.0917	0.0891	0.0792	0.0867	-0.52	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	162	0.0855	0.0883	0.1004	0.0914	-0.38	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	196	0.09	0.0914	0.0906	0.0907	-0.4	0.1046	14	0.0348	0
CBG (%AR)	010.01	AOAC 2018.10	183	0.0933	0.094	0.0939	0.0937	-0.31	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	200	0.095	0.097	0.096	0.096	-0.25	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	212	0.0971	0.1052	0.1094	0.104	-0.02	0.1046	14	0.0348	0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
CBG (%AR)	010.02	AOAC 2018.11, U	125	0.107	0.108	0.102	0.106	0.03	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	189	0.1146	0.1204	0.1198	0.118	0.39	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	122	0.1523	0.152	0.1521	0.152	1.37	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	155	0.214	0.19	0.201	0.202	2.79	0.1046	14	0.0348	0
CBG (%AR)	010.99	Other	106	0.25	0.25	0.26	0.253	4.28	0.1046	14	0.0348	0

CBGA (%AR)

CBGA (%AR)	011.02	AOAC 2018.11, U	125	<0.0018	<0.0016	<0.0020	<0.0018					1
CBGA (%AR)	011.99	Other	197	<0.0094	<0.0098	<0.0092	<0.0094					1
CBGA (%AR)	011.99	Other	216	<0.0099	<0.0091	<0.0103	<0.0099					1
CBGA (%AR)	011.99	Other	106	<0.0100	<0.0100	<0.0100	<0.0100					1
CBGA (%AR)	011.99	Other	162	<0.0100	<0.0100	<0.0100	<0.0100					1
CBGA (%AR)	011.99	Other	196	<0.0210	<0.0210	<0.0210	<0.0210					1
CBGA (%AR)	011.99	Other	194	<0.0270	<0.0270	<0.0270	<0.0270					1
CBGA (%AR)	011.99	Other	149	<0.0370	<0.0370	<0.0400	<0.0370					1
CBGA (%AR)	011.99	Other	122	<0.0500	<0.0500	<0.0500	<0.0500					1
CBGA (%AR)	011.01	AOAC 2018.10	183	0	0	0	0					2
CBGA (%AR)	011.99	Other	189	0.0245	0.0235	0.0236	0.0239					0

THCV (%AR)

THCV (%AR)	012.99	Other	197	<0.0094	<0.0098	<0.0092	<0.0094					1
THCV (%AR)	012.99	Other	216	<0.0099	<0.0091	<0.0103	<0.0099					1

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
THCV (%AR)	012.99	Other	106	<0.0100	<0.0100	<0.0100	<0.0100					1
THCV (%AR)	012.99	Other	196	<0.0210	<0.0210	<0.0210	<0.0210					1
THCV (%AR)	012.99	Other	194	<0.0270	<0.0270	<0.0270	<0.0270					1
THCV (%AR)	012.99	Other	149	<0.0370	<0.0370	<0.0400	<0.0370					1
THCV (%AR)	012.02	AOAC 2018.11, U	183	0	0	0	0					2
THCV (%AR)	012.99	Other	200	0.009	0.009	0.009	0.009					0
THCV (%AR)	012.99	Other	189	0.0142	0.0153	0.0157	0.0151					0

Δ8-THC (%AR)

Δ8-THC (%AR)	013.99	Other	162	<0.0100	<0.0100	<0.0100	<0.0100					1
Δ8-THC (%AR)	013.99	Other	200	<0.0100	<0.0100	<0.0100	<0.0100					1
Δ8-THC (%AR)	013.99	Other	196	<0.0210	<0.0210	<0.0210	<0.0210					1
Δ8-THC (%AR)	013.99	Other	194	<0.0270	<0.0270	<0.0270	<0.0270					1
Δ8-THC (%AR)	013.99	Other	149	<0.0370	<0.0370	<0.0400	<0.0370					1
Δ8-THC (%AR)	013.99	Other	102	<0.0484	<0.0188	<0.0187	<0.0484					1
Δ8-THC (%AR)	013.99	Other	183	0	0	0	0					2
Δ8-THC (%AR)	013.01	AOAC 2018.10	183	0.003	0.003	0.0031	0.003	-1.27	0.0209	6	0.0141	0
Δ8-THC (%AR)	013.99	Other	216	0.0187	0.0171	0.0189	0.0182	-0.19	0.0209	6	0.0141	0
Δ8-THC (%AR)	013.99	Other	106	0.02	0.02	0.02	0.02	-0.07	0.0209	6	0.0141	0
Δ8-THC (%AR)	013.99	Other	197	0.0194	0.0193	0.0206	0.0198	-0.08	0.0209	6	0.0141	0
Δ8-THC (%AR)	013.99	Other	189	0.0187	0.0251	0.0239	0.0226	0.12	0.0209	6	0.0141	0
Δ8-THC (%AR)	013.99	Other	138	0.0757	0.0687	0.0766	0.0737	3.74	0.0209	6	0.0141	0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
CBC (%AR)												
CBC (%AR)	014.99	Other	212	0.2289	0.2364	0.2337	0.233	-1.21	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	194	0.2436	0.2418	0.2174	0.234	-1.12	0.2501	14	0.0142	0
CBC (%AR)	014.01	AOAC 2018.10	183	0.2402	0.2354	0.2436	0.24	-0.73	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	106	0.25	0.25	0.22	0.24	-0.71	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	162	0.2395	0.2443	0.2405	0.241	-0.61	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	216	0.2404	0.246	0.2466	0.244	-0.41	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	149	0.243	0.245	0.248	0.245	-0.33	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	197	0.2479	0.2476	0.2566	0.251	0.04	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	196	0.2509	0.2532	0.2511	0.252	0.12	0.2501	14	0.0142	0
CBC (%AR)	014.02	AOAC 2018.11, U	125	0.251	0.256	0.252	0.253	0.21	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	155	0.265	0.274	0.257	0.265	1.08	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	189	0.2619	0.2673	0.2681	0.266	1.11	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	200	0.269	0.268	0.267	0.268	1.27	0.2501	14	0.0142	0
CBC (%AR)	014.99	Other	204	0.2667	0.2712	0.267	0.268	1.29	0.2501	14	0.0142	0
Cd (ug/g AR)												
Cd (ug/g AR)	101.04	ICP-MS	102	<0.008	<0.008	<0.008	<0.008					1
Cd (ug/g AR)	101.04	ICP-MS	194	<0.0200	<0.0200	<0.0200	<0.0200					1
Cd (ug/g AR)	101.04	ICP-MS	196	<0.0241	<0.0241	<0.0241	<0.0241					1
Cd (ug/g AR)	101.04	ICP-MS	106	<0.0750	<0.0750	<0.0750	<0.0750					1

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
Cd (ug/g AR)	101.04	ICP-MS	197	<0.0963	<0.1002	<0.0994	<0.0963					1
Cd (ug/g AR)	101.04	ICP-MS	162	<0.2000	<0.2000	<0.2000	<0.2000					1
As (ug/g AR)												
As (ug/g AR)	102.04	ICP-MS	102	<0.008	<0.008	<0.008	<0.008					1
As (ug/g AR)	102.04	ICP-MS	194	<0.0200	<0.0200	<0.0200	<0.0200					1
As (ug/g AR)	102.04	ICP-MS	196	<0.0482	<0.0482	<0.0482	<0.0482					1
As (ug/g AR)	102.04	ICP-MS	106	<0.0750	<0.0750	<0.0750	<0.0750					1
As (ug/g AR)	102.04	ICP-MS	197	<0.0963	<0.1002	<0.0994	<0.0963					1
As (ug/g AR)	102.04	ICP-MS	162	<0.2000	<0.2000	<0.2000	<0.2000					1
Hg (ug/g AR)												
Hg (ug/g AR)	103.04	ICP-MS	196	<0.0072	<0.0072	<0.0072	<0.0072					1
Hg (ug/g AR)	103.04	ICP-MS	106	<0.0080	<0.0080	<0.0080	<0.0080					1
Hg (ug/g AR)	103.04	ICP-MS	194	<0.0200	<0.0200	<0.0200	<0.0200					1
Hg (ug/g AR)	103.04	ICP-MS	197	<0.0963	<0.1002	<0.0994	<0.0963					1
Hg (ug/g AR)	103.04	ICP-MS	162	<0.2000	<0.2000	<0.2000	<0.2000					1
Hg (ug/g AR)	103.04	ICP-MS	102	0.0013	0.0014	0.0012	0.0013					0
Pb (ug/g AR)												
Pb (ug/g AR)	104.04	ICP-MS	102	<0.008	<0.008	<0.008	<0.008					1
Pb (ug/g AR)	104.04	ICP-MS	194	<0.0200	<0.0200	<0.0200	<0.0200					1

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
Pb (ug/g AR)	104.04	ICP-MS	196	<0.0241	<0.0241	<0.0241	<0.0241					1
Pb (ug/g AR)	104.04	ICP-MS	106	<0.0750	<0.0750	<0.0750	<0.0750					1
Pb (ug/g AR)	104.04	ICP-MS	197	<0.0963	<0.1002	<0.0994	<0.0963					1
Pb (ug/g AR)	104.04	ICP-MS	162	<0.2000	<0.2000	<0.2000	<0.2000					1
Cu (ug/g AR)												
Cu (ug/g AR)	105.04	ICP-MS	196	<0.4893	<0.4893	<0.4893	<0.4893					1
Ni (ug/g AR)												
Ni (ug/g AR)	106.04	ICP-MS	196	<0.0722	<0.0722	<0.0722	<0.0722					1
Cr (ug/g AR)												
Cr (ug/g AR)	107.04	ICP-MS	196	<0.0722	<0.0722	<0.0722	<0.0722					1
Cr (ug/g AR)	107.04	ICP-MS	197	<0.0963	<0.1002	<0.0994	<0.0963					1
Se (ug/g AR)												
Se (ug/g AR)	108.04	ICP-MS	197	<0.0963	<0.1002	<0.0994	<0.0963					1
Se (ug/g AR)	108.04	ICP-MS	196	<0.0963	<0.0963	<0.0963	<0.0963					1
Zn (ug/g AR)												
Zn (ug/g AR)	111.04	ICP-MS	196	1.84	1.66	1.89	1.8					0

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
Sb (ug/g AR)												
Sb (ug/g AR)	112.04	ICP-MS	196	<0.0241	<0.0241	<0.0241	<0.0241					1
(-)-alpha-Bisabolol (%AR)												
(-)-alpha-Bisabolol (%A	201.02	GC,MS	162	0.021	0.0165	0.0191	0.0189					0
(-)-alpha-Bisabolol (%A	201.02	GC,MS	196	0.1736	0.1655	0.16	0.166					0
(-)-alpha-Bisabolol (%A	201.02	GC,MS	194	0.1686	0.1852	0.1844	0.179					0
(-)-alpha-Bisabolol (%A	201.02	GC,MS	212	0.1906	0.2261	0.1899	0.202					0
beta-Caryophyllene (%AR)												
beta-Caryophyllene (%)	202.02	GC,MS	162	0.0819	0.0709	0.0768	0.0765					0
beta-Caryophyllene (%)	202.02	GC,MS	194	0.1333	0.1533	0.1302	0.139					0
beta-Caryophyllene (%)	202.02	GC,MS	212	0.1739	0.1965	0.1465	0.172					0
beta-Caryophyllene (%)	202.02	GC,MS	196	0.1817	0.1854	0.1718	0.18					0
(-)-Caryophyllene oxide (%AR)												
(-)-Caryophyllene oxide	203.02	GC,MS	194	0.0436	0.0419	0.0432	0.0429					0
(-)-Caryophyllene oxide	203.02	GC,MS	212	0.0487	0.0503	0.0449	0.048					0
(-)-Caryophyllene oxide	203.02	GC,MS	196	0.0689	0.0639	0.0635	0.0654					0
Geraniol (%AR)												
Geraniol (%AR)	204.02	GC,MS	194	<0.0004	<0.0004	<0.0004	<0.0004					1

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
Geraniol (%AR)	204.02	GC,MS	196	<0.0013	<0.0013	<0.0013	<0.0013					1
Geraniol (%AR)	204.02	GC,MS	162	<0.0100	<0.0100	<0.0100	<0.0100					1
alpha-Humulene (%AR)												
alpha-Humulene (%AR)	205.02	GC,MS	162	0.0225	0.0189	0.0202	0.0205					0
alpha-Humulene (%AR)	205.02	GC,MS	194	0.0495	0.0553	0.0489	0.0512					0
alpha-Humulene (%AR)	205.02	GC,MS	196	0.055	0.0549	0.0517	0.0539					0
alpha-Humulene (%AR)	205.02	GC,MS	212	0.0536	0.0653	0.0468	0.0552					0
d-Limonene (%AR)												
d-Limonene (%AR)	206.02	GC,MS	196	<0.0013	<0.0013	<0.0013	<0.0013					1
d-Limonene (%AR)	206.02	GC,MS	162	<0.0100	<0.0100	<0.0100	<0.0100					1
d-Limonene (%AR)	206.02	GC,MS	194	0.0025	0.0027	0.0025	0.0026					0
Linalool (%AR)												
Linalool (%AR)	207.02	GC,MS	162	<0.0100	<0.0100	<0.0100	<0.0100					1
Linalool (%AR)	207.02	GC,MS	212	0.0006	0.0013	0.0017	0.0012					0
Linalool (%AR)	207.02	GC,MS	196	0.0017	0.0016	0.0017	0.0017					0
Linalool (%AR)	207.02	GC,MS	194	0.0058	0.0058	0.0059	0.0058					0
beta-Myrcene (%AR)												
beta-Myrcene (%AR)	208.02	GC,MS	196	<0.0013	<0.0013	<0.0013	<0.0013					1

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
beta-Myrcene (%AR)	208.02	GC,MS	162	<0.0100	<0.0100	<0.0100	<0.0100					1
beta-Myrcene (%AR)	208.02	GC,MS	194	0.0023	0.0023	0.0024	0.0023					0
Ocimene (%AR)												
Ocimene (%AR)	210.02	GC,MS	194	<0.0004	<0.0004	<0.0004	<0.0004					1
Ocimene (%AR)	210.02	GC,MS	196	<0.0013	<0.0013	<0.0013	<0.0013					1
Ocimene (%AR)	210.02	GC,MS	162	<0.0100	<0.0100	<0.0100	<0.0100					1
alpha-Pinene (%AR)												
alpha-Pinene (%AR)	211.02	GC,MS	194	<0.0004	<0.0004	<0.0004	<0.0004					1
alpha-Pinene (%AR)	211.02	GC,MS	196	<0.0013	<0.0013	<0.0013	<0.0013					1
alpha-Pinene (%AR)	211.02	GC,MS	162	<0.0100	<0.0100	<0.0100	<0.0100					1
(-)-beta-Pinene (%AR)												
(-)-beta-Pinene (%AR)	212.02	GC,MS	194	<0.0004	<0.0004	<0.0004	<0.0004					1
(-)-beta-Pinene (%AR)	212.02	GC,MS	196	<0.0013	<0.0013	<0.0013	<0.0013					1
(-)-beta-Pinene (%AR)	212.02	GC,MS	162	<0.0100	<0.0100	<0.0100	<0.0100					1
alpha-Terpinene (%AR)												
alpha-Terpinene (%AR)	213.02	GC,MS	194	<0.0004	<0.0004	<0.0004	<0.0004					1
alpha-Terpinene (%AR)	213.02	GC,MS	196	<0.0013	<0.0013	<0.0013	<0.0013					1
alpha-Terpinene (%AR)	213.02	GC,MS	162	<0.0100	<0.0100	<0.0100	<0.0100					1

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
gamma-Terpinene (%AR)												
gamma-Terpinene (%A	214.02	GC,MS	194	<0.0004	<0.0004	<0.0004	<0.0004					1
gamma-Terpinene (%A	214.02	GC,MS	196	<0.0013	<0.0013	<0.0013	<0.0013					1
gamma-Terpinene (%A	214.02	GC,MS	162	<0.0100	<0.0100	<0.0100	<0.0100					1
alpha-Terpinolene (%AR)												
alpha-Terpinolene (%A	215.02	GC,MS	194	<0.0004	<0.0004	<0.0004	<0.0004					1
alpha-Terpinolene (%A	215.02	GC,MS	196	<0.0013	<0.0013	<0.0013	<0.0013					1
alpha-Terpinolene (%A	215.02	GC,MS	162	<0.0100	<0.0100	<0.0100	<0.0100					1
Δ9-THC (%DW)												
Δ9-THC (%DW)	501.01	AOAC 2018.10	210	0.091	0.091	0.089	0.0903					0
Δ9-THCA (%DW)												
Δ9-THCA (%DW)	502.01	AOAC 2018.10	210	<0.0460	<0.0430	<0.0470	<0.0460					1
CBD (%DW)												
CBD (%DW)	503.01	AOAC 2018.10	210	5.644	5.733	5.694	5.69					0
CBDA (%DW)												
CBDA (%DW)	504.01	AOAC 2018.10	210	<0.0460	<0.0430	<0.0470	<0.0460					1

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
Total Δ9-THC (%DW)												
Total Δ9-THC (%DW)	506.01	AOAC 2018.10	210	0.131	0.129	0.13	0.13					0
Total CBD (%DW)												
Total CBD (%DW)	507.01	AOAC 2018.10	210	5.684	5.771	5.735	5.73					0
CBDV (%DW)												
CBDV (%DW)	508.02	AOAC 2018.11, U	210	<0.0430	<0.0400	<0.0430	<0.0430					1
CBDVA (%DW)												
CBDVA (%DW)	509.01	AOAC 2018.10	210	<0.0460	<0.0430	<0.0460	<0.0460					1
CBG (%DW)												
CBG (%DW)	510.01	AOAC 2018.10	210	0.109	0.102	0.107	0.106					0
CBGA (%DW)												
CBGA (%DW)	511.01	AOAC 2018.10	210	<0.0460	<0.0430	<0.0460	<0.0460					1
THCV (%DW)												
THCV (%DW)	512.02	AOAC 2018.11, U	210	<0.0430	<0.0400	<0.0430	<0.0430					1

Analyte	Code	Method	Lab Num	Result1	Result2	Result3	Lab Value	Z score	Population of Lab Values			Flag
									Rob Mean	n	Rob StDev	
			Δ8-THC (%DW)									
Δ8-THC (%DW)	513.01	AOAC 2018.10	210	0.043	0.041	0.043	0.0423					0

1. Flag definitions: 1 (Lab Value not used because there were less than 2 numeric lab results), 2 (Lab Value not used because there were less than 2 numeric lab results greater than 0).

2. Robust statistics performed to determine robust mean and Z scores only if there were 6 or more numeric lab values.

3. AR refers to concentration on as-received basis. DW refers to concentration on dry weight basis.